



PROVINCIAL GOVERNMENT OF DAVAO DE ORO

Cabidanan, Nabunturan, Davao de Oro

TECHNICAL SPECIFICATIONS FORM

Purchase Request Number: 24-3368

Bid Number: B-24-0149
Date: 06/13/2024
Page: 1 of 1

ITEM NO.	QTY.	UNIT OF ISSUE	DESCRIPTION	APPROVED BUDGET		BID PRICES	
				Unit Price	Total Amount	Unit Price	Total Amount
1	1.0	pc	Autoclave 90-1580 Liters capacity with wire baskets	250,000.00	250,000.00		
2	1.0	pc	Biological Safety Cabinet Type II	316,264.00	316,264.00		
3	1.0	pc	Forced Air Drying oven Sterilizer at 170??c Class II, Type A, Wind Flow Type: Recirculation 70%, Exhaust 30%, Internal Dimension 1340 x 623 x 670 mm Approximately: Sterilization UV-20W; Stainless Exterior; Powder Coating with HEPA filter; High Glass Fiber Type Filter, 0.3 mcm; Particle 99.97% Cut-off or better; To provide stand or frame; High-efficiency double bower motor with double dustproof structure for low noise and low vibration	89,400.00	89,400.00		
4	1.0	pc	Incubator maintained at 35?c Interior; made of metal; capacity 85 L; Minimum Temp: 10 degrees Celsius; max 55 degrees Celcius	120,000.00	120,000.00		
5	1.0	pc	Water bath Incubator with wire basket temp at 44.5?c Metal Interior, with read-out digital thermometer; calibrated; preferably made in Germany	51,336.00	51,336.00		
TERMS AND CONDITIONS: On the Machine 1. The Supplier shall: a. Provide machine orientation and demonstration b. Conduct Monthly Preventive maintenance of the equipment consisting of cleaning and checking of equipment and monitoring and performing calibration of all the equipment. c. Provide 24/7 technical support. d. Provide the availability of technician and spear parts. e. Provide a maximum of two (2) years warranty.							
TOTAL/LOT					827,000.00		

Purpose: FOR PHO BLOOD USE 3RD QTR

Delivery: PGSO

Period of Delivery: 10 DAYS